1.   (Original)   A method for calibrating a
coincidence imaging system which includes a plurality of
radiation detectors, the method comprising:

measuring a plurality of coincidence radiation events
associated with a point radiation source;

assigning initial values for a set of fitting
parameters;

applying a minimization algorithm including:

calculating lines of response (LOR) based
upon   the   fitting   parameters   and   the   measured
radiation events,

generating a   figure of merit
characterizing the apparent size of the point
radiation source based upon the LOR's, and

optimizing the fitting parameters to
produce a minimized figure of merit; and

extracting from the optimized fitting parameters a
correction factor relating to a positional coordinate of a
detector.

2.   (Original)   A method for imaging using a
plurality of radiation detectors, the method comprising:

measuring a plurality of coincidence radiation events
associated with a point radiation source;

assigning initial values for at least one fitting
parameter;

calculating lines of response (LOR) based upon the
at least one fitting parameter and the measured radiation
events;

generating a figure of merit characterizing the
apparent size of the point radiation source based upon the
LOR's;

optimizing the at least one fitting parameter using
a minimization algorithm which includes iteratively repeating
the calculating and generating steps to produce a minimized
figure of merit;

extracting from the at least one optimized fitting parameter at least one correction factor;

acquiring a set of radiation data from an associated subject;

correcting the radiation data for camera misalignment by correcting the spatial coordinates of the detected radiation events using the at least one correction factor; and

reconstructing an image representation from the corrected radiation data.

3.　(Original)　The imaging method as described in claim 2, wherein the at least one fitting parameter includes:

a parameter related to the radial positional coordinate of a detector.

4.　(Original)　The imaging method as described in claim 2, wherein the at least one fitting parameter includes:

a parameter related to the tangential positional coordinate of a detector.

5.　(Original)　The imaging method as described in claim 2, wherein the at least one fitting parameter includes:

a parameter related to the orientational positional coordinate of a detector.

6.　(Original)　The imaging method as described in claim 2, wherein:

the step of generating a figure of merit includes summing a distance of closest approach of each LOR to a spatial point; and

the at least one fitting parameter includes the positional coordinates of the spatial point.

7.　(Original)　The imaging method as described in claim 2, wherein:

the step of generating a figure of merit includes summing the square of a distance of closest approach of each LOR to a spatial point; and

the at least one fitting parameter includes the positional coordinates of the spatial point.

8.   (Original) The imaging method as described in claim 7, wherein the step of generating a figure of merit further includes:

discarding LOR's whose distance of closest approach is greater than a preselected distance.

9.   (Original) The imaging method as described in claim 2, wherein the step of generating a figure of merit further includes:

obtaining a crossing point of each pair of LOR's; and

calculating a standard deviation of the crossing points.

10.   (Original) The imaging method as described in claim 2, wherein the step of generating a figure of merit further includes:

obtaining a distance of closest approach for each pair of LOR's; and

calculating a standard deviation of the obtained distances.

11.   (Original) The imaging method as described in claim 2, wherein the number of detectors is N and the fitting parameters include:

Δr_i, i=1 to N, where Δr_i is a correction for the radial coordinate of the ith detector;

Δt_j, j=1 to N, where Δt_j is a correction for the tangential coordinate of the jth detector; and

$\Delta\theta_k$, k=2 to N, where $\Delta\theta_k$ is a correction for the orientational coordinate of the kth detector.

12.   (Original) The imaging method as described in claim 11, wherein the fitting parameters further include:

positional coordinates of the point radiation source.

13.   (Currently Amended)   A method of PET imaging comprising:

coincidence detecting radiation events from a calibration source with at least two detector heads;

calculating correction factors that correct for mechanical misalignment of the detector heads from the coincidence detected calibration source radiation, the calculating including:

generating a figure of merit which characterizes an apparent size of a point source of radiation based on lines of response,

optimizing fitting parameters based on the figure of merit;

during a diagnostic imaging procedure performed on a subject, generating image data in response to radiation collected with the detector heads;

correcting the image data with the correction factors; and

reconstructing the corrected image data into an image representation.

14.   (Currently Amended)   A coincidence imaging system comprising:

a gantry;

a plurality of flat panel detectors disposed about the gantry;

a data memory which stores measured data about radiation events detected by the detectors;

a calibration memory which stores a plurality of calibration parameters for correcting data measured during a patient scan; and

a processor in communication with the calibration memory and with the data memory which calculates the calibration parameters by a minimization algorithm that includes: optimizing fitting parameters with respect to acquired radiation data associated with an apparent size of a point radiation source based on lines of response.

15.   (Currently Amended)   ~~The~~ A coincidence imaging system ~~of claim 14 wherein~~ comprising:

a gantry;

a plurality of detectors disposed about the gantry;

a data memory which stores measured data about radiation events detected by the detectors;

a calibration memory which stores a plurality of calibration parameters for correcting data measured during a patient scan; and

a processor in communication with the calibration memory and with the data memory which calculates the calibration parameters which are extracted from fitting parameters using a minimization algorithm, the minimization algorithm ~~further includes~~ including:

calculating lines of response (LOR) based upon the fitting parameters and the measured data;

generating a figure of merit characterizing the apparent size of the point radiation source based upon the LOR's; and

optimizing the fitting parameters to produce a minimized figure of merit.

16.   (Original)   The imaging system of claim 15 wherein the calibration parameters include:

parameters relating to positional coordinates of the plurality of detectors.

17.   (Original)   The imaging system of claim 16, wherein:

the gantry is a rotatable gantry which acquires measured data over a range of gantry angular positions.

18.   (Currently Amended)   The imaging system of claim ~~14~~ 15, wherein:

the figure of merit is generated by summing the square of a distance of closest approach of each LOR to a spatial point; and

the fitting parameters include the positional coordinates of the spatial point.

19. (Currently Amended) The imaging system of claim ~~14~~ 15, wherein the generating of the figure of merit includes:

obtaining a crossing point of each pair of LOR's; and

calculating a variance of the crossing points.

20. (Original) The imaging system of claim 14, wherein the minimization algorithm further includes:

discarding measured data about radiation events whose energy is outside a preselected energy range.